


WINNOCARE

F r a n c e

4 Le Pas du Château
85670 SAINT PAUL MONT PENIT
TEL : +33 (0)2 51 98 55 64
FAX : +33 (0)2 51 98 59 07
Email : info@medicatlantic.fr
Site Internet : <http://www.winnocare.fr>

USER MANUAL XXL XPRESS (XA1...) XXL DIVISYS (XB1...)



Creation 2017_01_01
Update 2020_05_01

CE

580166 Anglais



| | |
|----------------------------------------------------------------------------------------------------------------------------|-----------|
| 1. TRANSPORT AND STORAGE | 4 |
| 2. BED ENVIRONMENT CONDITIONS | 4 |
| 3. GENERAL USE | 4 |
| 3.1. PREREQUISITES BEFORE USE AND INSTRUCTIONS FOR USE | 4 |
| 3.1.1. TRAINING AND QUALIFICATIONS OF THE PERSON USING THE DEVICE | 4 |
| 3.1.2. INDICATIONS | 5 |
| 3.1.3. CONTRAINDICATIONS | 5 |
| 3.2. CLINICAL BENEFIT, DEVICE PERFORMANCE, ACTION MECHANISM DEPENDING ON THE FUNCTIONS, OPTIONS OR ACCESSORIES USED | 5 |
| 3.2.1. DEVICE PERFORMANCE CHARACTERISTICS | 5 |
| 3.2.2. EXPECTED CLINICAL BENEFITS | 5 |
| 3.2.3. INFORMATION FOR HEALTHCARE PROFESSIONALS | 5 |
| 3.3. PRECAUTIONS FOR USE | 5 |
| 3.4. ELECTRICAL CHARACTERISTICS | 7 |
| 3.4.1. ELECTRICAL DATA | 7 |
| 3.4.2. ELECTROMAGNETIC COMPATIBILITY | 8 |
| 3.4.3. EQUIPOTENTIALITY | 9 |
| 4. BED BOARDS 120/140/160CM WIDTH COMPATIBLE | 10 |
| 5. COMPATIBLE ACCESSORIES | 11 |
| 6. USE | 12 |
| 6.1. PURPOSE OF THE DEVICE | 12 |
| 6.2. SPECIFIC PRECAUTIONS FOR USE | 12 |
| 6.2.1. RESIDUAL RISKS AND ADVERSE SIDE EFFECTS | 12 |
| 6.3. GENERAL DESCRIPTION | 13 |
| 6.4. TECHNICAL CHARACTERISTICS | 13 |
| 6.4.1. DIMENSIONAL | 13 |
| 6.4.2. WEIGHT | 15 |
| 6.4.3. NOISE | 16 |
| 6.4.4. ELECTRICAL COMPONENTS | 16 |
| 6.4.5. PROTECTION LEVEL AGAINST DUST AND LIQUID PENETRATION | 16 |
| 6.5. ELECTRICAL CONNECTION DIAGRAM | 17 |
| 6.6. REMOTE CONTROL | 18 |
| 6.6.1. LOCKABLE REMOTE CONTROL HB02X (LOCK OR UNLOCK) | 18 |
| 6.7. BRAKING | 19 |
| 6.7.1. SEPARATE BRAKING | 19 |
| 6.7.2. CENTRALIZED BRAKING | 19 |
| 7. ASSEMBLY AND DISASSEMBLY | 20 |
| 7.1. ASSEMBLING THE BED XXL XPRESS | 20 |
| 7.2. MOUNTING XXL XPRESS BED AND XXL BED DIVISYS | 24 |
| 7.3. DISMANTLING | 27 |
| 8. OPERATION OF THE SLEEPING SURFACE | 29 |
| 8.1. BACK REST | 29 |
| 8.1.1. STANDARD BACK REST | 29 |

| | |
|----------------------------------------------------------------------------|-----------|
| 8.1.2. BACK REST WITH TRANSLATION | 29 |
| 8.1.3. EMERGENCY RELEASE OF THE BACK REST (CARDIO PULMONARY RESUSCITATION) | 30 |
| 8.2. LEG REST | 30 |
| 8.2.1. LEG REST WITH MANUAL CRANK (C/) | 30 |
| 8.2.2. LEG REST WITH ELECTRIC FOLDING (P/) | 30 |
| 9. INSTALLATION OF ACCESSORIES | 30 |
| <hr/> | |
| 9.1. BOARDS | 30 |
| 9.2. METAL SIDE RAILS | 31 |
| 9.3. WOODEN BARRIERS | 32 |
| 9.4. ALUMINUM HALF SIDE RAILS | 33 |
| 9.5. ANGLED LIFTING POLE AND IV STAND | 34 |
| 10. MAINTENANCE | 35 |
| <hr/> | |
| 10.1. IDENTIFICATION | 35 |
| 10.2. INSTRUCTIONS FOR DISMANTLING THE MOTORS | 35 |
| 10.3. MAINTENANCE | 36 |
| 10.4. QUALITY INSPECTION OF MEDICAL BEDS | 37 |
| 10.5. CLEANING AND DISINFECTION | 38 |
| 10.6. LIFETIME | 39 |
| 10.7. GARANTIES | 39 |
| 10.8. TROUBLESHOOTING GUIDE | 40 |
| 11. SCRAPPING | 41 |
| <hr/> | |

Dear Sir/Madam,

You have acquired a WINNCARE medical bed equipped with its accessories, and we thank you for your custom.

Our class I medical devices are designed and manufactured in accordance with Annex VIII of Regulation (EU) 2017/745.

They are tested in conformity with standard IEC 60601-2-52 (2010) in their commercial configurations, including the boards and accessories that we manufacture, so as to ensure you maximum safety and performance.

As a result, maintenance of the contracted good's warranty depends on compliance with the conditions for use recommended by WINNCARE and the use of original accessories, which also guarantees you safe use of the medical bed and its accessories.

1. TRANSPORT AND STORAGE

For transport, the bed should be in its low position, and strapped and protected. The wired control and supply lead should be attached to the bed base.

The head and footboards are protected and strapped to the sleeping surface.

The bed should be transported upright when in its original packaging in compliance with the instructions printed on the packaging.



It is strictly forbidden to stack packages weighing over 60kg/m², whatever position they are in.
Before transporting or dismantling the bed, make sure the back and leg rests are fixed to the frame of the bed base.


2. BED ENVIRONMENT CONDITIONS

The bed, along with the boards and accessories, must be transported and stored at a room temperature of between -10°C and +50°C,

The bed, along with the boards and accessories, must be used at a room temperature of between +10°C and +38°C,

Relative humidity of between 30% and 75%.

Atmospheric pressure between 700hPa and 1060hPa



Observe the specified environmental conditions

3. GENERAL USE

Before use, it is essential to read these instructions carefully. They contain advice on using and looking after the bed to guarantee optimum safety.

The latest updated version of this manual is available on our website www.winnicare.fr.

The photos, images and diagrams included in this manual may not correspond to your product depending on the versions or options you have.

3.1. Prerequisites before use and instructions for use

3.1.1. Training and qualifications of the person using the device

Users must be trained by people who have been trained and approved by the relevant business entities, in particular when it comes to safety and reporting non-conformities.



The user and staff must be trained and aware of the risks associated with using the bed. He must not allow it by children and be vigilant when used by confused or disoriented people.

3.1.2. Indications

Suitable for patients with temporary or permanent decreased independence due to their illness or disability.

3.1.3. Contraindications

To be determined by the prescribing doctor based on the patient's condition, the medical treatment and the type of relevant accessories. Patient weight higher than the safe operating load defined.

3.2. Clinical benefit, device performance, action mechanism depending on the functions, options or accessories used

3.2.1. Device performance characteristics

Variable height (ergonomics of care, transfer aid), low bed (reduces the severity of falls from the bed), Trendelenburg, reverse Trendelenburg, chest rest, leg rest, flat, chair position (preserves or improves the patient's physical and/or psychological state, helps with administering care), 30° chest rest and/or with Fowler or semi-Fowler positions (helps prevent pressure sores), 5° Trendelenburg (helps lift the patient), rails (protects the patient and/or caregivers), lateral stop (stabilizes the mattress), removable bed panels (prevents the patient from falling out of bed, acts as a fixed walking aid, helps with head and foot care).

3.2.2. Expected clinical benefits

To preserve the patient's state of health, mobility, activity and transfers. To reduce the repercussions of the immobilization syndrome. To protect the health and safety of users or others.

3.2.3. Information for healthcare professionals

Identify possible combinations of risks and benefits for each intended use. Assess the patient's ability to use and understand the overhead remote control and the caregivers' knowledge of how the medical bed and its accessories work.

Any serious incident occurring in connection with the device must be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is established. Inform the competent authority if you think or have a reason to believe that the device presents a serious risk or is faulty.

3.3. Precautions for use

Although the bed is conforming with Electromagnetic Compatibility, some devices may alter how it functions, in which case they must be used at a distance or not used at all.

The bed is a medical device and must not be modified under any circumstances. You must ensure its traceability, including that of the boards and its accessories.

If you are assembling medical devices not provided by the bed manufacturer, you must check the conformity of the assembly and make the EU declaration of conformity of the new medical device.

The electric parts (jack, supply box, wired control, etc.) shall only be repaired by the manufacturer Linak.

The bed is not suitable for use with an inflammable anaesthetic mixture with air or oxygen or nitrous oxide.

The loads permitted (see bed characteristics) must be distributed evenly over the bed base.

Do not activate all the motors at the same time when the patient is in the bed (only one motor is authorised at one time, except elevation by 2 motors or simultaneous function).

After each use and while care is being administered to the patient, the brakes must be activated. We recommend putting the bed in its low position after every use and while the patient is resting, to reduce the height of falls by a confused or agitated person. Remember to lock the function(s) (if the option is available).

On change of height or angle of the parts of the bed, make sure that there are no objects and no parts of the patient's or carer's body caught between the bed, the boards, the accessories and the ground or between the boards and base or between the cross braces.

Do not sit down on the side of the back rest or leg rest if this is not flat. In the case of a prolonged more than 50 ° tilt bust semi-sitting position, it is recommended to vary the position of the person in bed every 2 hours.

When the bed is being moved, keep the power lead well away from the ground and wheels.

When use of an adaptor, extension lead or connection plug proves necessary, you must check that its characteristics are suitable for the bed.

Connection to the supply box must be done using a mains complying with the standards in force and corresponding to a voltage of use indicated on the device.

The mains plug must be disconnected before the bed is moved.

Do not pull on the mains leads to disconnect the mains plug.

During any handling, try not to catch the leads of the motors and remote control and do not get them knotted.

The wired control must be hooked to the headboard when not in use. In the case of the use of infrared remote control(s), WINNCARE allows the establishment of a single bed in the same room (or in a close environment) or a second bed only if the infrared options of 2 beds concerned are different (I **and** I1).

The condition of the leads must be checked frequently. If the slightest modification is observed, the person in charge for maintaining the bed must be contacted to carry out the necessary repairs.

If repairs are required, the person in charge of maintenance must be contacted.

For greater safety, some side rails can be adapted (see accessories). Side rails should not be used to manipulate or move the medical bed. To assist patient mobility, it is possible to fit a Mobility Aid System (S.A.M.TM).

For assistance, if necessary, in mounting, operation or maintenance or to report unexpected operation or events, call your supplier or Winnicare. The cleaning instructions recommended must be complied with.

Only use original parts and accessories supplied by WINNCARE to guarantee safety and maintain product conformity. The bed must not be modified.

Abnormal use of the bed may damage it or cause accidents to users, in which case the warranty shall be annulled. Abnormal use means failure to comply with the precautions for use, maintenance instructions and other uses not related to the bed's normal purpose, such as: use of the bed by several people at the same time (except DUO DIVISYS bed), use outdoors, moving the bed on a slope that is steeper than 10°, etc.

Put the bed in the designated room, foreseeing an appropriate perimeter of use for the different functions (variable height, TR, etc.), especially if the bed has a lifting pole or side rails. Check that there is sufficient ceiling height if a lifting pole is fitted.

Brake the wheels.

The mains socket should remain accessible to enable the bed to be disconnected quickly.

Plug in the power lead, checking that the mains comply with the standards in force and that it is suitable for the supply box voltage.


Also ensure that the power lead, the remote control lead as well as the cables of possible other devices are positioned correctly to prevent any risks of getting caught between the moving parts of the bed.



- Check that the bed operates properly after installing it in accordance with the check-list appended in this document. (Test all of its functions)
- The patient is a planned operator of the bed. Users must be trained in how to use the equipment.
- Inform the patient and his visitors of the safety instructions to be observed.

3.4. Electrical characteristics

3.4.1. Electrical data



3.4.2. Electromagnetic compatibility



The bed is an electromedical device requiring special precautions with regard to electromagnetic compatibility. The device must be installed and put into service according to the EMC information provided in this manual.



The use of accessories, transducers, and cables other than those specified or supplied by the manufacturer may cause increased electromagnetic emissions or a decrease in the immunity of the device and may cause improper operation.

The bed will not move automatically when subject to electromagnetic disturbances within the limit of the values indicated below.

| Manufacturer's declaration and guide – electromagnetic emissions | | |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| The medical bed (see references in contents) has been designed for use in the electromagnetic environment specified below. The user should ensure that it is used in such an environment. | | |
| Emissions test | Compliance | ELECTROMAGNETIC ENVIRONMENT - GUIDE |
| RF emissions CISPR 11 | Group 1 | The medical bed (see references in contents) uses RF energy only for its internal functions. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. |
| RF emissions CISPR 11 | Class B | The medical bed (see references in contents) can be used in all domestic environments, including those directly connected to the public low-voltage power supply network that supplies buildings for domestic purpose. [] |
| Harmonic emissions EN 61000-3-2 | Class A | |
| Voltage fluctuations / Flicker EN 61000-3-3 | Compliant | |
| RF emissions CISPR 14-1 | Not Applicable | The medical bed (see references in contents) has not been designed for connection to other equipment. |
| Manufacturer's declaration and guide - electromagnetic immunity | | |
| The medical bed (see references in contents) has been designed for use in the electromagnetic environment specified below. The user should ensure that it is used in such an environment. | | |
| IMMUNITY TEST | IEC 60601 Severity level | COMPLIANCE LEVEL |
| Electrostatic discharge EN 61000-4-2 | ± 8 kV contact ± 15 kV air | ± 8 kV contact ± 15 kV air |
| Radiated RF EN 61000-4-3 | 3 V/m 80Mhz – 2,7Ghz 80% AM at 1kHz | 3 V/m 80Mhz – 2,7Ghz 80% AM at 1kHz |
| Proximity fields from RF wireless communication equipment EN 610004-3 | See table below | See table below |
| Electrical fast transients EN 61000-4-4 | ± 2 kV for feeders ±1 kV for input/output lines Repetition frequency at 100 kHz | ± 2 kV for feeders ±1 kV for input/output lines Repetition frequency at 100 kHz |
| Surges EN 61000-4-5 | Differential mode ± 1 kV Common mode ± 2 kV | Differential mode ± 1 kV Common mode ± 2 kV |
| Conducted RF EN 61000-4-6 | 3V 0,15 – 80 MHz 6V in ISM band between 0.15 and 80 MHz | 3V 0,15 – 80 MHz 6V in ISM band between 0.15 and 80 MHz |
| Magnetic fields IEC 61000-4-8 | 30A/m 50 Hz or 60 Hz | 30A/m 50 Hz or 60 Hz |
| Voltage dips EN 61000-4-11 | 0% U _T ; 0,5 cycle at 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315° 0% U _T ; 1 cycle 70% U _T ; 25/30 cycles at 0° | 0% U _T ; 0,5 cycle at 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315° 0% U _T ; 1 cycle 70% U _T ; 25/30 cycles at 0° |
| Voltage interruptions EN 61000-4-11 | 0% U _T ; 250/300 cycles | 0% U _T ; 250/300 cycles |
| NB: U _T is the nominal value of power voltage applied during the test. | | |


| Test specification for immunity to RF wireless communications devices | | |
|-----------------------------------------------------------------------|----------------------------------|---------------------------|
| Test frequency (MHz) | Modulation | Immunity test level (V/m) |
| 385 | Pulse Modulation : 18 Hz | 27 |
| 750 | FM + 5 Hz deviation : 1 kHz sine | 28 |
| 710 – 745 – 780 | Pulse Modulation : 217 Hz | 9 |
| 810 – 870 – 930 | Pulse Modulation : 18 Hz | 28 |
| 1720 – 1845 – 1970 | Pulse Modulation : 217 Hz | 28 |
| 2450 | Pulse Modulation : 217 Hz | 28 |
| 5240 – 5500 - 5785 | Pulse Modulation : 217 Hz | 9 |

3.4.3. Equipotentiality

Under the head-half of the bed base you will find an equipotentiality socket, identified by the label, enabling you to connect any electromedical devices. The leads of these devices must pass through the head end and not on the side.



Equipotentiality




4. BED BOARDS 120/140/160cm WIDTH COMPATIBLE

| Item | Width 120 |  KG | Width 140 |  KG | Width 160 |  KG |
|------------------------------------|-----------|--------------------------------------------------------------------------------------|-----------|----------------------------------------------------------------------------------------|-----------|----------------------------------------------------------------------------------------|
| MEDIDOM II (2) | P319-00 | 25 | P320-00 | 28 | P321-00 | 32 |
| WINNEA | P322-00 | 25 | - | - | - | - |
| MEDIDOM II ELEGANT (2) | - | - | P324-00 | 30 | P325-00 | 33 |
| Louis Philippe (1)(2) | P416-00 | 21 | P417-00 | 27 | P418-00 | 29 |
| Auzence II | P612-00 | 22 | P613-00 | 25 | P614-00 | 28 |
| Abélia II | P617-00 | 19 | P618-00 | 22 | P619-00 | 25 |
| CARMEN II with wooden barriers (2) | P621-00 | 38 | P622-00 | 43 | P623-00 | 49 |
| CARMEN II | P626-00 | 34 | P627-00 | 38 | P628-00 | 43 |
| ABELIA II Aluminum side rails (1) | P640-00 | 25 | P643-00 | 29 | - | - |
| STYLVIA Aluminum side rails (1) | P641-00 | 40 | P644-00 | 45 | - | - |
| NOVIDA Aluminum side rails (1) | P642-00 | 40 | P645-00 | 45 | - | - |

- (1) Bed board incompatible with the XPRESS transport kit
 (2) Long pan option incompatible with the XPRESS transport kit

➤ Mattress



Observe the mattress dimensions prescribed. See user guide

| Width of base in cm | Characteristics of compatible mattresses |
|---------------------|----------------------------------------------------------------------------------------------------------------------------------|
| 120 | Width 116 cm minimum with a high resilience foam of 34Kg/m³ minimum |
| 140 | Width 136 cm minimum with a high resilience foam of 34Kg/m³ minimum |
| 140 DUO | Width 68 cm minimum with a high resilience foam of 27Kg/m³ minimum |
| 160 DUO | Width 78 cm minimum with a high resilience foam of 27Kg/m³ minimum |
| 160 DUO option V | 1x Width 68 cm minimum + 1x Width 88 cm minimum , with a high resilience foam of 27Kg/m³ minimum |



Incompatible mattresses can pose RISKS.



There must be at least 220 mm between the top of the side rail and uncompressed and no therapeutic mattress surface. It will be advisable to tend towards this specification in the case of the use of a therapeutic mattress.

5. COMPATIBLE ACCESSORIES



Only accessories and boards supplied by **WINNCARE** guarantee you a compatibility with our products.



Incompatible accessories and boards can pose RISKS

| Ref. | Des. (x1 : incompatible with bed base extension) | Art.* | max (CFS)** | AERYS 120 | DIVISYS 120 | DIVISYS 140 | XPRESS 120 | XPRESS 140 | DUO 140 | DUO 160 |
|-----------------|-----------------------------------------------------|-------|--------------------|-----------|-------------|-------------|------------|------------|---------|---------|
| | | | | x | x | x | x | x | x | x |
| A1700xx | IV stand, 2 hooks | 0.6 | 8 | x | x | x | x | x | x | x |
| A8400xx | Telescopic IV stand, 2 hooks | 1.7 | 8 | x | x | x | x | x | x | x |
| A193-00 | Chrome-plated wall-mounted basin holder | 0.3 | NA | x | x | x | x | x | x | x |
| A230-00 | Remote-control lead holder | 1.1 | NA | x | x | x | x | x | x | x |
| A260-00 | Epoxy urinal holder | 0.2 | NA | | x | x | x | x | x | x |
| A551-00 | Excelys wall stop | 1.1 | NA | | x | x | | | | |
| A553/555-00 | Skirt for 120/140 and L.P beds | 16.4 | NA | | x | x | | | | |
| A554/556-00 | Skirt for 160 and L.P beds | 17.6 | NA | | | | | | x | x |
| A575-00 | Support handle | 1.3 | NA | x | x | x | x | x | x | x |
| A579/580/581-00 | Wooden side rails (1) | 10.5 | NA | x1 | x1 | x1 | x1 | x1 | x1 | x1 |
| A676-00 | | | | | | | | | | |
| A613-00 | S.A.M. BLOCK | 1.7 | NA | x | x | x | x | x | x | x |
| A616/617-00 | S.A.M. ACTIV | 7.4 | NA | x | x | x | x | x | x | x |
| A622-00 | Angled lifting pole | 4.6 | 75 | x | x | x | x | x | x | x |
| A623-00 | Chromed angled lifting pole | 4.7 | 75 | x | x | x | x | x | x | x |
| A631-00 | Bed base extension L140 cm | 9.4 | NA | | | x | | x | x | |
| A632-00 | Bed base extension L160 cm | 10.3 | NA | | | | | | | x |
| A637/638-00 | EPOXY side rail | 11.6 | NA | x | x | x | x | x | x | x |
| A639/640-00 | Chrome barrier | 11.8 | NA | x | x | x | x | x | x | x |
| A645/646-00 | S.A.M. EVOLUTION | 20.6 | NA | x | x | x | x | x | x | x |
| A647/648-00 | S.A.M. ERGONOM PLUS | 23.6 | NA | x | x | x | x | x | x | x |
| A649/650-00 | S.A.M. ERGONOM | 13.6 | NA | x | x | x | x | x | x | x |
| A651/652-00 | S.A.M. ERGONOM LIGHT | 11.4 | NA | x | x | x | x | x | x | x |
| A653-00 | Kit for side loading bed Aerys | 9.3 | NA | x | | | | | | |
| A660-00 | Bed base extension L120 cm | 7.8 | NA | x | x | | x | | | |
| A663-00 | Aerys 120 wall stop | 1.2 | NA | x | | | | | | |
| A665/666-00 | Aluminum half side rails | 31.6 | NA | x | x | x | x | x | x | x |
| A673-00 | Long universal pan | 4.6 | NA | x1 | x1 | x1 | x1 | x1 | x1 | x1 |
| S0200 | Stainless steel intravenous stand on base | 2.5 | 8 | x | x | x | x | x | x | x |
| Y0300 | Lifting pole on U shaped base, fixed height | 23.3 | 75 | x | x | x | x | x | x | x |

* Item

** Normal load in use

6. USE

6.1. Purpose of the device

- The device is intended to compensate for a handicap and as a preventive measure.
- XXL X'Press & Divisys beds, depending on their configurations, are intended for home treatment (application environment 4).
- XXL X'Press & Divisys beds are intended for persons whose size is greater than or equal to 146 cm, with a weight greater than or equal to 40 kg and a body mass index (BMI) greater than or equal to 17.

Warning refer to user's manual

Calculation of BMI :

$$\text{BMI} = m / h^2$$

m : weight in kg
h : size in m

6.2. Specific precautions for use


The bed should not be used as a stretcher.
Moving the bed with weight requires 2 people.

The kickstand and transport kit can be removed only if the bed is placed on the ground on its wheels.
It is imperative to raise the kickstand on the ship kit before transporting the bed.

6.2.1. Residual risks and adverse side effects

| DANGER | RESULT | REQUIREMENTS FOR THE USER |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------|------------------------------------------------------------------------------------------------------|
| Trapping | Hands pinched | Handle legsrest or backrest up by handles or at specified locations |
| Trapping | Crushing body or object | Before lowering the bed, check that the area between the bed and the floor is free |
| Trapping | Crushing body or object | Check before operating the central brake pedal that the area between the pedal and the floor is free |
| ADVERSE SIDE EFFECTS | | |
| Fall of the person. Immobilization syndrome. Aggravation of fractures that have not been stabilized and/or procedural pain when the articulated parts of the bed frame are moved. | | |



6.3. General description






① Manual crank or electric folding leg rest depending on version.

6.4. Technical characteristics

6.4.1. Dimensional







DIVISYS BED WITH CENTRALIZED BRAKING OPTION

DIVISYS BED WITH OPTION SKIRT ON FOOT

DIVISYS BED WITH ACCESSORY SKIRT ON WHEELS

6.4.2. Weight

Safe operational load of the bed: 315 kg (Patient 270 kg, Mattress 30 kg, Accessories 15 kg).

| | <i>Base and brace</i> | <i>Head-end bed base</i> | <i>Foot-end bed base</i> | <i>TOTAL</i> |
|-----------------------------------------------|-----------------------|--------------------------|--------------------------|--------------|
| <i>Bed 120 XPRESS (C/)</i> | 55 kg | 25 kg | 29 kg | 109 kg |
| <i>Bed, version manual crank leg rest</i> | | | | |
| <i>Bed 120 XPRESS (P/)</i> | 55 kg | 25 kg | 35 kg | 115 kg |
| <i>Bed, version electric folding leg rest</i> | | | | |
| <i>Bed 140 XPRESS (C/)</i> | 55 kg | 28 kg | 31 kg | 114 kg |
| <i>Bed, version manual crank leg rest</i> | | | | |
| <i>Bed 140 XPRESS (P/)</i> | 55 kg | 28 kg | 39 kg | 122 kg |
| <i>Bed, version electric folding leg rest</i> | | | | |
| <i>Bed 120 DIVISYS (C/)</i> | 47kg | 25 kg | 28 kg | 100 kg |
| <i>Bed, version manual crank leg rest</i> | | | | |
| <i>Bed 120 DIVISYS (P/)</i> | 47 kg | 25 kg | 34 kg | 106 kg |
| <i>Bed, version electric folding leg rest</i> | | | | |
| <i>Bed 140 DIVISYS (C/)</i> | 47 kg | 28 kg | 30 kg | 105 kg |
| <i>Bed, version manual crank leg rest</i> | | | | |
| <i>Bed 140 DIVISYS (P/)</i> | 47 kg | 28 kg | 38 kg | 113 kg |
| <i>Bed, version electric folding leg rest</i> | | | | |

6.4.3. Noise

The measurement of the maximum audible sound power in accordance with ISO 3746 is 49 dB(A).

6.4.4. Electrical components

| DESIGNATION | TYPE | PROTECTION INDEX | V/Hz |
|------------------------------------|------|------------------|-------------------------|
| Absorbed power | | 200VA | |
| Control box | CA40 | IP66 | 100-240 VAC 50-60 Hz |
| Control box | CB6 | IP66 | 230 VAC 50 Hz |
| Actuator | LA27 | IP66 | 24 VDC |
| Actuator | LA40 | IP66 | 24 VDC |
| Wired control | HB7x | IP66 | 24 VDC |
| Lockable and backlit wired control | HB02 | IP66 | 24 VDC |
| Lockable wired control | HL7x | IP54 | 24 VDC |
| Infrared control | HB23 | IP21* | 3 VDC |



* IP21 = Keep dry





Maximum operating time: Read the recommendations on the electrical label on the bed.

6.4.5. Protection level against dust and liquid penetration



| Index | 1 st number (decade) Protection against solids | 2 nd number (unit) Protection against water intrusion |
|-------|--------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------|
| 0 | No protection. | No protection. |
| 1 | Protected against solid bodies greater than 50 mm. | Protected against vertical drops of water drops. |
| 2 | Protected against solid bodies greater than 12,5 mm. | Protected against falling drops of water up to 15° from the vertical. |
| 3 | Protected against solid bodies greater than 2,5 mm. | Protected against rain water up to 60 ° from vertical. |
| 4 | Protected against solid bodies greater than 1 mm. | Protected against splashing water from all directions. |
| 5 | Protected against dust and other microscopic residues. | Protected against jets of water from all directions at the lance (6.3 mm nozzle, distance 2.5 m to 3 m, flow 12.5 l / min ± 5%). |
| 6 | Totally protected against dust. | Protected against strong jets of water from all directions to the lance (12.5 mm nozzle, distance 2.5 m to 3 m, flow 100 l / min ± 5%). |

6.5. Electrical connection diagram



6.6. Remote control

Carry out a test cycle when the bed is empty to familiarise yourself with the bed functions.




6.6.1. Lockable remote control HB02x (lock or unlock)




6.7. Braking

6.7.1. Separate braking

Check that the wheels are locked by trying to move the bed. If this is not done, the patient or another person who leans on the bed may fall.





6.7.2. Centralized braking







7. ASSEMBLY AND DISASSEMBLY

7.1. Assembling the bed XXL XPress




Lower the holders




1

Remove the foot from the bed







2

Remove head from the bed






3











7.2. Mounting XXL Xpress bed and XXL bed DIVISYS






9






7.3. Dismantling

Repeat the operations in the reverse order 10; 9; 8; 7; 6 and 5





Finish by repeating operations 2 and 1.in the reverse order.


8. OPERATION OF THE SLEEPING SURFACE

8.1. Back rest


8.1.1. Standard back rest

In the event of a power cut or failure, flatten the back rest as follows :

- a) Disconnect the power supply.
- b) Dismantle the headboard ①.
- c) Stand at the head of the bed and take hold of the back rest handle ② with one hand. Push or lift to compensate the pressure exerted by the patient and unhook the clip ③ by the connecting rod with the other hand. The back rest jack will then pivot downwards.
- d) Put the headboard back.



Version with handle on the back rest




Version without handle on the back rest

8.1.2. Back rest with translation

Same operation for flattening as the standard back rest.

The back rest with translation avoids to slip forward..



8.1.3. Emergency release of the back rest (Cardio Pulmonary Resuscitation)



- 1) Grasp the back rest with one hand.
- 2) With the other hand, activate one of the two handles on the back rest while lowering.

If the handle is released, the back rest will stop moving.



Release handles



To reactivate the cardiopulmonary resuscitation system of high-load beds (270Kg), actuate the backrest electric actuator to its minimum position by pressing the down button on the remote control.

8.2. Leg rest

8.2.1. Leg rest with manual crank (C/)

To lift, lift the leg rest using the handle at the end.

To lower, relieve the leg rest slightly or to its maximum with the hand to release the catch, then lower the leg rest.

8.2.2. Leg rest with electric folding (P/)

Memory folding: This function keeps a position of the tibia section horizontal when the jack is activated upwards. To use this function, the 1st crank catch must be engaged when the leg rest is flat.

Folding without memory : the end of the tibia section stays in contact with the bed base.


9. INSTALLATION OF ACCESSORIES

9.1. Boards

- Installing a bed board :




➤ Removing a bed board :



9.2. Metal side rails

➤ To positioning the side rail :



① Insert amount of the side rail folded in the grip jaw on the head side of the bed to hear “CLIC”.


② Tighten the Rondo screws of the grip jaws on the bed base.

To remove the side rail folded, loosen the rondo on the feet side ② and then release the amount ① without removing, pulling the index finger on the side of the jaw on the head side. Take the side rail, a hand at each end to remove it completely.




If the side rail is poorly positioned, safety of the patient may be endangered or a malfunctioning may occur. The side rails must not be used when the patient is a child (under 12) or if s/he is too small (≤ 146 cm).

- To lower the side rail :



- To raise the side rail :



Check that the side rail is locked by trying to fold it without using the release pin

9.3. Wooden barriers


See the attached instructions for fitting the wooden side rail.

- To raise the side rail.


- ① Raise the top side rail with both hands until it locks.
- ② Check that it is properly slot in.

- To lower the side rail.

- ① Raise the top side rail with one hand.
- ② Press on the unlocking button with the other hand.
- ③ Support the rail as it lowers.
- ④ Repeat steps ① to ③ for the other side.



Check that the high guide engage in the right direction in the low guide.




9.4. Aluminum half side rails

See the attached instructions for fitting the aluminum half side rails.


➤ To raise the side rail.

1-In intermediate position :

① Raise the top side rail with both hands until it locks.




② Check that it is properly slot in by trying to get it down without using the unlocking buttons.




2-In high position :

① Raise the top side rail with both hands to the top by pushing the unlocking buttons then release them before the high stop.




② Check that the side rail is locked by trying to fold it without using the unlocking buttons.



➤ To lower the side rail.

- ① Raise the top side rail.
- ② Press on the two unlocking buttons with both hands.
- ③ Support the rail as it lowers.



9.5. Angled lifting pole and IV stand

In case of psycho-behavioral problems of the patient, it is necessary to fix the stem to the bed with the screw and the washer provided.




10. MAINTENANCE



No maintenance should be done in the presence of a patient on the bed.

10.1. Identification




10.2. Instructions for dismantling the motors



Unplug the device before dismantling.

- Dismantle when the bed is empty or in the side position.
- If dismantling in any other position, keep a firm hold of the moving parts to avoid any shearing.
- Unblock the safety clips ①, unplug the motor leads, and remove them from the securing seals.
- Put the motors back in place and put in the same direction as at the beginning.



10.3. Maintenance

Quality control of medical beds will be made by technical staff or trained biomedical and taking into account the normal conditions of use specified in the user guide, on a bed equipped with its accessories, including its barriers if necessary.

The bed must be available to perform all quality control at least once a year, but also on special request and corrective maintenance on the performance that could be affected by the intervention. However, to save time this may be associated with preventive maintenance. In this case, it is not useful to make a further examination of already controlled performance.

RECOMMENDATIONS FOR PREVENTIVE MAINTENANCE:

Preventive maintenance should be carried out in accordance with our specifications and at least once a year by the organisation or person who installed the bed.

Between two maintenance sessions and at least once a year, the following should be carried out:

- Verification that the electrical leads are connected all along the metal jambs to prevent shearing of these leads when the variable height is being activated.
- Verification that all of the electrical leads and plugs are in good condition. Replacement if there is the slightest alteration (wear, shearing, damage, etc.).
- Verification of the external appearance (traces of damp and good overall condition of protective covers in particular) and that the motors and jacks function properly.
- Verification that the bed is in good working order (test all functions).
- Verification that the frame, bed base and mechanical joints are all in good condition.

When maintenance is carried out at the patient's home as part of a long-term contract, the installer must also:

- Check that the bed is properly installed (check to see that there hasn't been any modification contrary to the safety instructions made by the user since the bed's installation).
- Remind the users of the safety instructions.
- All installation and preventive maintenance operations must be recorded. See table model below. This record must be kept in a designated area throughout the bed's lifetime.

10.4. Quality inspection of medical beds

ANNUAL MAINTENANCE OF MEDICAL BEDS

IDENTIFICATION

ESTABLISHMENT

| | | | | | |
|--------------------------|---|---|---|---|---|
| Type model / Trademark : | | | | | |
| Date of manufacture : | | | | | |
| Serial or park number : | | | | | |
| Application Environment | 1 | 2 | 3 | 4 | 5 |

VISUAL CONTROL

| | Not applicable | Congruent | Improper |
|------------------------------------------------------------------|-----------------------|-----------------------|-----------------------|
| Identification - label | | <input type="radio"/> | <input type="radio"/> |
| General cleanliness | | <input type="radio"/> | <input type="radio"/> |
| Condition and attachment of head and footboards | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |
| Lifting pole (fastening, strap, winder, ...) | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |
| Sleeping surface (bedspring) | | <input type="radio"/> | <input type="radio"/> |
| Verification of tightenings, diverse nuts and bolts, pins, pivot | | <input type="radio"/> | <input type="radio"/> |
| State of corrosion, welding | | <input type="radio"/> | <input type="radio"/> |
| Remote controls (displays, lights and locking) | | <input type="radio"/> | <input type="radio"/> |
| Bed specific side rails | | | |
| Space between head of bed and siderail (< 60 mm)* | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |
| Space between the half siderails (< 60 ou > 318 mm)* | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |
| Condition, mounting direction, adaptation to bed ... | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |

FONCTIONAL CONTROL

| | Not applicable | Congruent | Improper |
|---------------------------------------------------------------------|-----------------------|-----------------------|-----------------------|
| Mobile parts (variable height, backrest, legrest, Trendelenburg...) | | <input type="radio"/> | <input type="radio"/> |
| Remote controls including functions locking | | <input type="radio"/> | <input type="radio"/> |
| Castors (pivoting, rolling) | | <input type="radio"/> | <input type="radio"/> |
| Braking | | <input type="radio"/> | <input type="radio"/> |
| Fixing of siderails and locking in high position (closed) | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |
| Absence of sound disturbances (squeaking, lubrications) | | <input type="radio"/> | <input type="radio"/> |

CPR CONTROL

| | Not applicable | Congruent | Improper |
|---------------------------------------------------------------|-----------------------|-----------------------|-----------------------|
| Retracting headboard in case of emergency (App. envir. 1 & 2) | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |
| Emergency flattening of backrest | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |

ELECTRICAL CONTROL

| | Not applicable | Congruent | Improper |
|----------------------------------------------------------------------------------|-----------------------|-----------------------|-----------------------|
| Status of electrical cables, plugs and connectors | | <input type="radio"/> | <input type="radio"/> |
| Status of electrical equipment (controlbox, actuators,...) | | <input type="radio"/> | <input type="radio"/> |
| Recurrent test following standard NF EN 62353 (see test report for more details) | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |

Remains in service : YES YES subject to the processing of observations NO

Obsolescence : *New*  *Decrepit*

COMMENTS

| | | |
|-----------|-----------------------------------|-----------------------|
| Made on : | | |
| Name : | Signature and stamp of technician | Signature of customer |

*following standard IEC- 60601-2-52

10.5. Cleaning and disinfection

High pressure cleaning, machine cleaning, washing tunnel or jet cleaning is prohibited.
Unplug the mains lead.

Check that all the electrical parts are connected together. All the sockets of the supply box must be used or filled, otherwise its watertightness is not guaranteed.

Clean the electric covers of the jacks and wired control by means of a mop impregnated with a detergent-disinfectant product.

The medical bed is a non-critical appliance requiring "Low Level" disinfection.

We draw your attention to the fact that the recommendations below are drawn up according to the rules of good practice but are not a protocol. Contact the hospital hygiene department.

➤ AIM

To recondition the bed and prevent the transmission of germs from one patient to another.

To eliminate all organic soiling by:

- physical action (cleaning)
- chemical action (disinfection)

➤ INDICATION

Physical and bacteriological cleanliness of the bed and its accessories

➤ EQUIPMENT

Microfiber wipes

Detergent or Detergent-Disinfectant (Surface DD with CE mark) and surface Disinfectant (Chlorine substances, alcohol base < 30%)

Attention: DD products and Javel water must not be used undiluted. A remanence time must be applied according to the disinfectant manufacturer's instructions (the drying time without human presence is often the same as the remanence time) (Refer to the instruction manual)

SANIVAP steam appliance with accessories

➤ TECHNICAL

- Daily maintenance with a surface DD product in one operation.
- Maintenance when the patient leaves, or periodically, by the process known as *Bio cleaning* observing the 3 operations:
 - Cleaning is done by means of a cloth soaked in a surface detergent or Detergent-Disinfectant (DD) solution
 - Rinsing is done with cloth rinsed in clean water
 - Disinfection is done by means of a cloth soaked in a surface disinfectant solution.
- Specific maintenance by specialist contractors after removal of the bed from the establishment:
 - Dispose of the packaging after decontamination of the inside by spraying with a Detergent-Disinfectant solution
 - *Bio cleaning* operation, or,
 - Steam cleaning (accessory with microfiber band) of the flat surfaces and the base slats. Change the washing mops regularly to prevent water accumulating. Clean the parts that are difficult to access with a steam nozzle (wheels, hinges after opening, corners, etc.). For tubes, use the steam nozzle with a microfiber cloth. Never direct the nozzles onto electrical boxes or actuators.
 - Dry hinges with compressed air
 - Attention: Disinfect jacks, electrical boxes and remote controls with a microfiber cloth soaked in disinfectant.

Do not rinse or wipe.

- Check the operation of all the bed functions
- Repair if necessary
- Pack in thermoplastic film

Warning

- In the event of additional precautions (Contact precautions, Droplets or Air), apply the measures recommended by the hospital hygiene department
- The use of a Javel water solution of more than 5000ppm (0.5% of active chlorine) should be justified by a microbiological risk and only applied for the required time (Risk of ageing of some materials, especially their colour).
- The concentration of alcohol-based surface disinfectant solutions should be less than 30%.

Note: The use of the terminal disinfection process is compatible with the medical bed and its accessories.



Product for external use. Do not swallow, keep away from heat sources and avoid contact with eyes.

10.6. Lifetime

The lifetime under normal conditions of use and maintenance is 8 years for beds and 5 years for accessories and boards.

10.7. Garanties

- All of our products carry a warranty against any manufacturing defect, provided the normal conditions for use and maintenance are complied with.
- Labour costs due to changes in structures or parts under warranty are not taken into account.
- Please refer to the standard terms of sale for the specific terms of warranty for each product.
- Every time you contact us for possible maintenance, you must quote us the information on the bed identification label and on the electric parts if these are concerned.
- Original parts shall be supplied for replacement, within the term of warranty, by our customer sales network determining the beginning of the term of warranty.
- Defective parts must be returned to ensure proper application of this warranty and also to avoid any invoicing.

10.8. Troubleshooting Guide

| DEFAULT | POSSIBLE CASE | HELP |
|-----------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|
| The whole bed is not working | The bed is not electrically powered | Check the electrical connection (Green power LED on controlbox) |
| | The remote control is fully locked | Unlock the remote control |
| | The infrared remote control is not powered or is not connected to the receiver. | Remove the rear cover Check batteries |
| | | Reverse the small contactor on top of the remote control |
| | One or more cables are not connected properly | Check the correct connection of the electrical cables Check for the presence of anti-tear clips at the cable connections |
| | One or more cables are cut or pinched | Check the condition and passage of electrical cables |
| An electrical element is defective | Contact after sales service | |
| The entire bed does not work, a short and discontinuous beep sounds when a key is pressed on the remote control | The controlbox is faulty | Do a reset of the controlbox (§ Specific use) or contact after sales service if a reset is impossible |
| A continuous beep sounds when a key is pressed on the remote control | The backup battery has reached its last cycle | Reconnect the bed to the mains to recharge the battery |
| A long and discontinuous beep is emitted when the bed is in the high position (normal event) | The bed is no longer braked (position on wheels, normal event) | Press down on the variable height to remove the wheels (normal use) |
| The selected function of the bed does not work | The function is locked on the remote control | Unlock the function on the remote control |
| | The movement has reached the stop | Operate the opposite direction of movement on the remote control |
| | An external element blocks the mechanism | Remove the blocking element |
| | One or more cables are not connected properly | Check the correct connection of the electrical cables Check for the presence of anti-tear clips at the cable connections |
| | | One or more cables are cut or pinched |
| | An electrical element is defective | Contact after sales service |
| The selected function of the bed is interrupted during movement | The movement has reached the stop | Operate the opposite direction of movement on the remote control |
| | The movement has reached a memorized position (normal event) | Release the button on the remote control and then press (normal cycle) |
| | An external element blocks the mechanism | Remove the blocking element |
| Bed can not be moved / moves with difficulty | The putting on wheels is not carried out or incomplete | Press the remote control to put the bed on its wheels until the movement stops |
| | Directional wheel hinders movement | Put the pedal in the 4-wheel free position |
| | An external element blocks the mechanism | Remove the blocking element |
| | One or more wheels are defective | Contact after sales service |

11. SCRAPPING

The product must be scrapped if the main requirements are no longer met, particularly when the product no longer has its original characteristics and has not been subject to corrective action during the manufacturing process.

Measures should therefore be taken to ensure that the bed is no longer used for the purpose it was originally intended.



The crossed-out bin symbol indicates that the electrical elements of the Winncare product are considered to be Electrical and Electronic Equipment (EEE), subject to EU Directive 2012/19 / EU on the Selective sorting of electrical and electronic waste (DEEE); batteries and accumulators that can be used in these products fall within the scope of the European Union and Council Directive 2006/66 / EC of 6 September 2006 on batteries and accumulators and waste batteries and accumulators. The complete product and / or the easily detachable electrical part and the batteries / accumulators contained in Winncare products must therefore be separately sorted in the countries of the European Union.

WINCARE

F r a n c e

4 Le Pas du Château

85670 SAINT PAUL MONT PENIT

TEL : +33 (0)2.51.98.55.64

FAX : +33 (0)2.51.98.59.07

Email : info@medicatlantic.fr

Site Internet : <http://www.winncare.fr>